

Certificate of Conformity

VWR Disposable Laboratory Coat, SPP, L

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 113-8299

Lot No: 27120217

Colour: White

Material: 45gsm (SPP)

Date of Manufacture: 10th Sep 2021

Quality System Compliance: Products are manufactured under the ISO13485:2016 standard. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

QC testing:

Inspected according to AQL level II 4.0 sampling plan per MIL-STD-105E

Break Tensile properties(N): ISO 13934-1 standard, MD = 75.42, CD=45.47

Elongation rate (%): ISO 13934-1 standard, MD = 79.45, CD=77.91

BSE/TSE Statement: The above product contains no ingredients of animal origin and no material derived from, or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform Encephalopathy (TSE)/Bovine Spongiform Encephalopathy (BSE)

Other Characteristics: Latex Free

This document has been produced electronically and is valid without a signature

Date of Issue: **Feb 2022**