

Corning Incorporated - Life Sciences
271 County Route 64
Big Flats NY 14814 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: FUNNEL, 75MM, PP, 12/12	Manufacture Date	: 2019-01-29
Catalog Number	: 6120P-75		
Lot ID	: 02919001		
Expiration Date	: Unconfirmed Validity		

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass
Packaging Inspection - Pass
Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



ANDREW J (ANDY) FRANKS
CMO Quality Manager