
Certificate of Conformity

Product: Nunc Immuno F16 Module, MaxiSorp,
Flourescence, Black

Date: 2021-06-01

Page: 1

Item number: 475515

Lot number: 170922

Date of manufacture: April 2021

Date of expiry: April 2026

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
Binding Variation, tested by adsorption of IgG	Plate CV% less than 10 %	Complies

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality management system - Requirements" and ISO 13485:2016 "Medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site



Thomas Andersen
Operations Quality Manager

Please note that this is an electronically generated signature