



Certificate of Conformance and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	PSME0001
Lot Number:	1087810-0-0-1
Manufacturing Date:	08/28/2024
Isopropyl Alcohol:	70 ± 3% by volume with the remainder DI H2O.
Expiration Date:	08/2026

Sterility

Gamma Processing Run ID:	3617608
Irradiation Dose Specified:	25.0 kGy to 50.0 kGy
Irradiation Dose Delivered:	29.1 kGy to 39.7 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)

A handwritten signature in black ink, appearing to read "C. Duffee".

QC Supervisor
coc@contecinc.com

Tuesday, September 10, 2024

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec Europe
ZI Du Prat, Avenue Paul Duplex
56000 VANNES France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com

Page 1 of 1

DID: 8621B



Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
 TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

09/09/2024 13:07:36 GMT
 Page 1 of 1

Customer Name:	Contec- PO Box 530	Processing Facility:	Charlotte Gamma	Work Order #	3617608
P.O.#	143793			Sales Order #	3631687
25-50 kGy		CNT6, Gamma Treatment	Irradiation Date/Time:	09/06/2024 11:26:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	60	CA	PSME0001	Clean Room Product	1080185-0-0-1	143793
102.000	60	CA	PSME0001	Clean Room Product	1087810-0-0-1	143793
103.000	60	CA	PSME0001	Clean Room Product	1088186-0-0-1	143793
104.000	72	CA	PSCS0009	Clean Room Product	1068674-0-0-1	143793
105.000	72	CA	PSCS0009	Clean Room Product	1068675-0-0-1	143793
106.000	216	CA	MS16IPAST	Clean Room Product	1083983-0-0-1	143793
107.000	216	CA	SB167030IR	16 oz IPA Spray Bottle 70/30	1081811-0-0-1	143793
108.000	216	CA	SB167030IR	16 oz IPA Spray Bottle 70/30	1081812-0-0-1	143793
	972	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By ----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	29.1 KGY	Pass	BTOWERY Benjamin Towery	09/06/2024 23:57:19 GMT
450.00	Maximum Dose	25.0 kGy	50.0 kGy	39.7 KGY	Pass	BTOWERY Benjamin Towery	09/06/2024 23:57:41 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: BRITTANY LONG
 Reason: Work Order Completions

Date: 09/09/2024 13:06:07 GMT