

PRODUCT CERTIFICATE
Certificate of Compliance

September 24, 2021

Catalog Number: **8060-0200**
Lot Number: **27699957**
Description: **Nalgene 50 Platinum-Cured Silicone Tubing**
Manufacturing Tolerances: **Close**
Date of Manufacture: **01/14/2021**

Product Material:

- Meets or exceeds FDA Regulations CFR 177.2600 for contact with food.
- Meets or exceeds USP Class VI Certification.
- Complies with European Pharmacopoeia, Section 3.1.9.
- Meets or exceeds ISO 10993 Testing Requirements

Thermo Fisher Scientific certifies this material supplied has been manufactured with Nalgene 50 Brand Platinum-cured Silicone Elastomer in conformance with cGMP guidelines and meets or exceeds the following:

Material Analysis:

Property Tested	Specification Range	
	Minimum	Maximum
Shore A	50	56
Modulus @100%	120	N/A
Tensile PSI	1,000	N/A
Elongation	700	N/A
Tear B PPI	200	N/A

Product defined above was produced and inspected in the USA, in accordance and in compliance with established manufacturing procedures and quality acceptance requirements.

This material is not intended for implantation.

Sincerely,

Tammy Parmelee

Tammy Parmelee
Quality Systems Administrator
Thermo Fisher Scientific
Labware and Specialty Plastics
CertificateROC@thermofisher.com

Rev. 1