Acetone
ChromAR HPLC

Material No: 2435-25
Effective from: 2013/06/29

Product Specification

Test	Specification
Assay (GC, corrected for water)	>= 99.8 %
Water (KF)	<= 0.2 %
Titrable Acid (meq/g)	<= 0.0005
Titrable Base (meq/g)	<= 0.0006
Residue after Evaporation	<= 0.0005 %
Ultraviolet Absorbance (1.00-cm cell vs. water):	
330 nm	<= 0.8
340 nm	<= 0.07
350 nm	<= 0.01
355 nm	<= 0.004
UV Cut-off	<= 327 nm

For Laboratory, Research or Manufacturing use.
Suitable for Liquid Chromatography and UV-Spectrophotometry

Barbara Rykowska
Quality Assurance Coordinator

For questions on this Product Specification please contact Technical Service at technical.service@avantorsciences.com
Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel: +48 322392000