

Certificate of Conformity

VWR Sampling Spoon, 50 ml, White

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

Cat. No: 231-0490

Lot.No.: 190503MEI53789901

Expiry date: Apr 2021

Material: Polystyrene

Length : 161 mm

Type of Irradiation : Gamma

Date of Irradiation : 17th May 2019

Required Minimum Dose : 10 kGy

Determined Minimum Dose : 20.9 kGy

Determined Maximum Dose : 31.5 kGy

This meets the requirements of FDA 21 CFR part 11.

This document has been produced electronically and is valid without a signature

Date of issue: **September 2019**