



## Quality Certificate

### VWR® MICROCENTRIFUGE TUBES, POLYPROPYLENE

As per the manufacturer, the below product meets the following criteria:

|                                   |                                         |
|-----------------------------------|-----------------------------------------|
| <b>North American Catalog No:</b> | 87003-294                               |
| <b>Lot Number:</b>                | 639927-D23401                           |
| <b>Description:</b>               | VWR Tube Microcent Natural 1.7 MI Pk500 |
| <b>Expiration Date:</b>           | September 2023                          |

**Quality Management System** – Complies with the current version of the EN ISO 13485:2016 Standard.

**BSE/TSE** – Product complies with the latest revision of EMA/410/01 “*Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products*” by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

**Non-Pyrogenic** – Tested and met the criteria established in the current version of ANSI/AAMI ST 72, “*Bacterial Endotoxins – Test methodologies, routine monitoring, and alternatives to batch testing*”. The acceptance level for product is  $\leq 0.02$  EU/ml or  $\leq 0.8$  EU/device.

**USP Class VI Testing** – All material resin is tested, qualified and shown to be non-toxic as established in the Standard USP Class VI Chapter <87>, “*Biological Reactivity Test, in Vitro*” and Chapter <88>, “*Biological Reactivity Tests, in vivo*.”

**Human and Bacterial DNA Free** – Tested by PCR method and found to be free of detectable human and bacterial DNA contamination. The assay detection limit is hDNA 0.5pg/ul and bDNA 0.0001pg/ul.

**PCR Inhibition** – Tested with real time Polymerase Chain Reaction and found no inhibition of real time PCR with a Ct comparison of  $\leq 5XCT_{PosCtrl-Ct}$ .

**DNase/RNase Free** – Tested by nuclease assay and found to be free of detectable DNase/RNase contamination. The assay detection limit is  $4 \times 10^{-6}$  U/ul for DNase and  $10^{-10}$  U/ul for RNase.

**Sterility** – Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137-3:2017, “*Sterilization of health care products- Radiation*”. Products meet a minimum Sterility Assurance Level (SAL) of  $10^{-3}$ .

**Quality Control Testing** – Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspections and inline tests are:

- Visual Inspection – Pass
- Packaging Inspection – Pass
- Fit Form Function Testing – Pass

The product met the manufacturer's high standards of quality at the time of batch/lot release.

Signed:

Jamie Ethier  
VP Global Quality  
VWR, Part of Avantor

Date: October 15, 2020