

YANGZHOU MEDLINE INDUSTRY CO., LTD.

No. 108, Jinshan Road, Economic Development Zone, Yangzhou, China

To whom it may concern,

CERTIFICATE OF CONFORMITY

The products listed as enclosed had been manufactured according with ISO 13485:2016 and China GMP requirements.

(pls see list as enclosed)

I. Visual Inspection: conform;

II. Functional Inspection: Conform;

III. Sterilization: Conform;

IV. LAL Test (Absence of bacterial endotoxins): Conform;

The visual and functional inspections were completed in accordance with Terumo specifications based on EN/ISO standards.

Sterilization method: EO Sterilization

Sterilization parameters are determined and revalidated periodically according to EN ISO 11135-1.

Sterility Assurance Level (SAL) is 10^{-6} .

The absence of bacterial endotoxins was tested in accordance with USP specifications.

YANGZHOU MEDLINE INDUSTRY CO., LTD.



YANGZHOU MEDLINE INDUSTRY CO., LTD.

No. 108, Jinshan Road, Economic Development Zone, Yangzhou, China

2011-TE-3/2101-TE-1

Product code	Lot No	Sterilization Date	Exp. Date
MDSS01SE	20210308T	2021-03-08	2026-03-07
MDSS01SE	20210309T	2021-03-09	2026-03-08
MDSS01S2516E	20210402T	2021-04-02	2026-04-01
MDSS01S2613E	20210403T	2021-04-03	2026-04-02
MDSS01SE	20210404T	2021-04-04	2026-04-03
MDSS05SE	20210407T	2021-04-07	2026-04-06
MDSS20ESE	20210409T	2021-04-09	2026-04-08

