



AquaPhoenix[®]
S C I E N T I F I C



CHEMetrics[®]
By AquaPhoenix Scientific

Simplicity in Water Analysis

Certificate of Conformance

Catalog No: K-7540

Product Description: Oxygen Kit, Dissolved

Lot: 4A144

Date of Manufacture: 24 May 2024

Date Certified: 24 May 2024

Expiration Date: N/A

Test	Specification	Answer
Reagent Appearance	colorless	Pass
Dissolved Oxygen Test at 0 ppb	0 to 5 ppb	Pass
Dissolved Oxygen Test at 80 ppb	60 to 100 ppb	Pass

CHEMetrics by AquaPhoenix Scientific, LLC hereby certifies that the following water analysis products have been manufactured and evaluated in accordance with strict Quality Control (QC) and Quality Assurance (QA) protocols and/or conform to manufacturer specifications. Where applicable, laboratory test results as well as retention samples for each production batch are on file at CHEMetrics.

Jennifer R. Nash

Signature

Jennifer R. Nash

Director of Quality



AquaPhoenix[®]
S C I E N T I F I C



CHEMetrics[®]
By AquaPhoenix Scientific

Simplicity in Water Analysis

Certificate of Conformance

Catalog No: C-7540

Product Description: Oxygen Comparator (round)

Lot: 4GD1209

Date of Manufacture: 26 April 2024

Date Certified: 26 April 2024

Expiration Date: April 2027

Test	Specification	Answer
Reagent Appearance	gradient from colorless to pale pink then to dark pink	Pass

CHEMetrics by AquaPhoenix Scientific, LLC hereby certifies that the following water analysis products have been manufactured and evaluated in accordance with strict Quality Control (QC) and Quality Assurance (QA) protocols and/or conform to manufacturer specifications. Where applicable, laboratory test results as well as retention samples for each production batch are on file at CHEMetrics.

Jennifer R. Nash

Signature

Jennifer R. Nash
Director of Quality



AquaPhoenix[®]
S C I E N T I F I C



CHEMetrics[®]
By AquaPhoenix Scientific

Simplicity in Water Analysis

Certificate of Conformance

Catalog No: R-7540

Product Description: Oxygen Refill, Dissolved

Lot: 4GE0600

Date of Manufacture: 16 May 2024

Date Certified: 16 May 2024

Expiration Date: August 2026

Test	Specification	Answer
Reagent Appearance	colorless	Pass
Dissolved Oxygen Test at 0 ppb	0 to 5 ppb	Pass
Dissolved Oxygen Test at 80 ppb	60 to 100 ppb	Pass

CHEMetrics by AquaPhoenix Scientific, LLC hereby certifies that the following water analysis products have been manufactured and evaluated in accordance with strict Quality Control (QC) and Quality Assurance (QA) protocols and/or conform to manufacturer specifications. Where applicable, laboratory test results as well as retention samples for each production batch are on file at CHEMetrics.

Jennifer R. Nash

Signature

Jennifer R. Nash
Director of Quality