

Certificate of Conformity

VWR Catalogue Number	689633-010101
Description	Petri dishes, PS, with 3 vents, 90 mm Diameter, 15 mm Height, Sterile
Lot Number	040124
Date of Manufacture (yyyy-mm-dd)	2024-04-01
Date of Sterilization (yyyy-mm-dd)	2024-04
Expiry Date (yyyy-mm-dd)	2029-04-01
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2024-05-21

Quality System Compliance

Products are manufactured under the **ISO 9001:2015** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Polystyrene (PS), Color: Transparent.
Sterilization	Product is ETO treated and they meet a minimum requirement of 10^{-6} Sterility assurance level. Sterilisation procedure is validated in conformity with ISO 11135.
ATP Assay	Not Applicable.
DNase & RNase Free	Not Applicable.

BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not Applicable.
Non-Pyrogenic Statement	Not Applicable.
Latex Statement	Not Applicable.
BPA Statement	Bisphenol are not used in the manufacture of the raw material and are not expected to be present.
DEHP Statement	Not Applicable.
RoHS	No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr6+), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Bis(2-Ethyhexyl) Phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analyse is performed).
REACH Statement	Not Applicable.
Storage Conditions	Not Applicable.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.

Annex

Height of Bottom:	13.20 – 13.80 mm
Diam of top bottom:	87.00 – 88.00 mm
Diam of bottom:	86.50 – 87.10 mm
Height of lid:	7.60 – 8.20 mm
Diam of lid:	90.9 – 91.5 mm