



## CERTIFICATE OF MANUFACTURE/ANALYSIS

|                   |                         |               |                     |
|-------------------|-------------------------|---------------|---------------------|
| Product Size:     | <u>5 Gallons</u>        | Product Name: | <u>LabKlenz 100</u> |
| Manufacture Date: | <u>October 10, 2018</u> | Product ID:   | <u>1L10-05</u>      |
| Exp. Date:        | <u>N/A</u>              | Lot No.:      | <u>296641</u>       |

I. The product listed above is made per the current issue of the following documentation:

Raw Material Quality Control Specifications  
Finished Product Master Formula  
Manufacturing Instructions  
Packaging Specifications  
Finished Product Quality Control Specifications

Copies of this documentation are on file at STERIS Corporation, St. Louis Operations and will be provided as required to the applicable regulatory agencies.

II. The product listed above has been analyzed. The release specifications and results are shown below.

| <u>Test Name</u>                        | <u>Method No.</u> | <u>Results</u> | <u>Specification</u>      |
|-----------------------------------------|-------------------|----------------|---------------------------|
| Color                                   | 110               | Light Yellow   | Colorless to Light Yellow |
| Clarity                                 | 107               | Clear          | Clear                     |
| pH (1% w/w soln)                        | 106               | 11.5           | 11.0 – 12.0               |
| Total Alkalinity (as Na <sub>2</sub> O) | 4                 | 5.0%           | 4.5 – 5.5% w/w            |
| Specific Gravity (@ 25°C)               | 114               | 1.155          | 1.145 – 1.165             |

Based on these results the product is released for shipment to all STERIS Corporation distribution points.

Signed:

Date: October 17, 2018

Quality Representative