



Material	83609.600
Material description	Buffer and Checking solution Set
Grade	AVS TITRINORM

Characteristics	Specifications
Titer (KCl)	2.95 --> 3.05 mol/l
Electrode cleaning solution	Passes test
pH 4 (20°C)	3.98 --> 4.02
pH 7 (20°C)	6.98 --> 7.02
pH 10 (20°C)	9.95 --> 10.05

Signature

We certify that this batch conforms to the specifications listed above.

This document has been produced electronically and is valid without a signature.

Ellen Somers, Head of Laboratory - Haasrode
 VWR International bvba; Geldenaaksebaan 464; BE-3001
 Leuven; Belgium



For Professional use in Laboratory or Manufacturing. Not for use as an Active Pharmaceutical Ingredient or Food or Animal Feed. Suitability and intended use of the product remains the responsibility of the user.