

Certificate of Conformity

VWR Test Tube Rack

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 212-0425

Lot No: V39-260819

No. Of Holes	For tube Ø (mm)	Colour	WxDxH (mm)	Pk
24 (3x8)	30	Blue	110×285.5×86.7	1

Material: 30 % Glass filled Nylon with color masterbatch.

Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2015 & ISO 13485:2016 Quality Management System. Inspection records are reviewed and signed for product release.

QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols

Test Name	Test Reference	Accuracy	Result
Autoclaving	QWI-QC-07	General Insp. Level-1, AQL -1.5	PASS
Dimensional	QSF-QC-23	General Insp. Level-1, AQL -1.5	Within Tolerance

Certificate of Conformity

BSE/TSE Statement: The above mentioned product is manufactured without the use of Raw materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our knowledge.

Latex Free Statement: No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

This document has been produced electronically and is valid without a signature

Date of issue: **March 2021**