

CERTIFICATE OF CONFORMITY & CERTIFICATE OF ANALYSIS

Brand	MICROFLEX®
Size(s)	M (7.5 - 8)
Material	Nitrile
Style	MidKnight® MK-296
Short Description	Distinctive, Black Nitrile Exam Glove
Lot Number	2412LC
Batch Number	24120151LC
Manufacturing Date	2024-12
Expiration Date	2027-12

CERTIFICATE OF CONFORMITY

Ansell certifies that the above product is manufactured in accordance with Ansell product specifications and is in conformity with the following standards or normative documents:

Standard/Normative Document	Result
FDA 21 CFR 177-2600	Conforms
Reach Regulation	Conforms

CERTIFICATE OF ANALYSIS

Property	Test Method	Specification	Result
Freedom From Holes (General Inspection Level I)			
Freedom from Holes - Final Release	EN455-1/ASTM D5151	G1, AQL 1.5	Conforms
Visual Inspection			
Visual Inspection-major	Internal test method	G-I, AQL 2.5	Conforms
Visual Inspection-minor	Internal test method	G1, AQL 4.0	Conforms
Physical Properties			
Material Standard	ASTM D6319	S2, AQL 4.0	Conforms
Dimension Test	ASTM D6319/ASTM D3767	S2, AQL 4.0	Conforms
Finger Thickness (mm)	ASTM D6319/ASTM D3767	0.12- 0.19	Conforms
Palm Thickness (mm)	ASTM D6319/ASTM D3767	0.09 - 0.13	Conforms
Cuff Thickness (mm)	ASTM D6319/ASTM D3767	0.09 -0.15	Conforms
Glove Length (mm)	ASTM D6319/ASTM D3767	Min. 240	Conforms
Palm Width (mm)	ASTM D6319/ASTM D3767	72-130	Conforms
Physical Properties	ASTM D412	S2, AQL 4.0	Conforms
Tensile Strength (MPa)	ASTM D412	Min. 16	Conforms
Ultimate Elongation (%)	ASTM D412	Min. 400	Conforms
Modulus at 300% elongation (MPa)	ASTM D412	Max. 9	Conforms
Additional Properties			

Property	Test Method	Specification	Result
Residual Powder Content (mg/glove)	ASTM D6124	Max. 2.0	Conforms
Protein Content (µg/g)	ASTM D5712	N/A	Conforms

Electronically Signed by:

Sandra Wessel-Jess

QA Manager - Americas

Date: 29 July 2025

For additional information visit us at www.ansell.com, or contact us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.
Riverside Business Park
Blvd International, 55
B-1070 Brussels,
Belgium
T: +32 (0) 2 528 74 00
F: +32 (0) 2 528 74 01

Australia & Asia Pacific Region

678 Victoria St,
Richmond VIC 3121
Australia
Tel: +61 1800 337 041
Fax: +611 800 803578
Email: anzcustomerservice@ansell.com

North America Region

Ansell Healthcare products LLC
111 S Wood Ave,
Suite 900 Iselin,
NJ 08830,
United States
T: +1 800 800 0444
F: +1 800 800 0445
CA Tel: +1 800 363 8340
Email: customerserviceus@ansell.com

Ansell Healthcare Products LLC
2301 Robb Drive,
Reno,
NV 89523,
United States
US Tel: +1 800 800 0444
US Fax: +1 800 800 0445
Email: customerserviceus@ansell.com

Latin America & Caribbean Region

Ansell Comercial México S.A. de C.V.
Blvd. Bernardo Quintana 7001-C,
Q7001 Torre 2, Suites 1304-1306 Centro Sur.
Querétaro, Qro. México C.P. 76090
Tel: +52 442 296 2050
Email: cslac@ansell.com

Ansell Brazil Ltda.

Rua das Figueiras, 474
4º andar - Bairro Jardim
Santo André - SP
CEP: 09080-300 – Brasil
Tel: +55 11 3356 3100
Email: sac@ansell.com