

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Monterrey, NL, MEX.



Ruben Deschamps
 Sr. Quality Manager
 04/24/2026

The following information represents Product Certification for: Item#: **3662-0025**

Description: **GRADUATED CYLINDR PPCO; 25ML**

Lot#: **7468938010**

Manufactured: **01/22/2026**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-3205-48	BASE,25ML,GRAD,CYL,PP,BLUE	COMPONENT PART				
8-0029-14P	RESIN,PPCO,BLU,INJ	COLOR MIX (PPCO, BLU, RESIN)	N/A	N/A	N/A	N/A
8-0029-04	RESIN,PP,COPOLYMER,INJ	POLYPROPYLENE COPOLYMER	1698	N/A	N/A	177.1520(a)(3)(i) and(c)3.1a, 3.2a
8-0099-74	COLOR,BLUE,MULTI	COLORANT, BLUE, MULTI-RESIN	N/A	PASSED	PASSED	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210
1-3205-68	BODY,25ML,GRAD,CYL,PP,MOLDED WIP	COMPONENT PART				
8-0028-05	RESIN,PP,COPOLYMER,INJ	POLYPROPYLENE COPOLYMER	7478	N/A	PASSED	177.1520(a)(3)(i), (c)3.1a,except for cooking, (useconditions C-H)

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

NALGENE Graduated Cylinders are designed to meet accuracy requirements of ASTM Class B, E1272, "Cylinder, Graduated, Laboratory, Glass" and all requirements of ISO Standard 6706, "Plastic Laboratory Ware- Graduated Measuring Cylinders."

ANIMAL DERIVED MATERIALS (BSE/TSE) - All resins and colorants used at the manufacturing site comply with the latest revision of EMA/410/01 section 6.4.