

Certificate of Conformity

VWR Transport Swab Shaft/Head, Transport Medium, Plastic/Viscose, Amies Gel

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 710-0432

Lot No: 201923000

Expiry Date: 30th June 2022

Sterility: Products labelled "STERILE R" have been gamma irradiated.
Sterilization procedure is validated in conformity with BS EN ISO 11137-1:2015 and BS EN ISO 11137-2:2015.

Goods manufactured according to ISO 13485:2016 system

Medium Visual Appearance and Mechanical Controls: **passed according to internal SOP.**

pH of the medium at 25 °C acceptability limits 7.2 ± 0.3 : **passed**

Microscopic Bioburden Specifications, maximum 2 micro-organisms per 20 microscopic fields at 10X100 magnification: **passed.**

Cultural Response: swabs were inoculated with the tester organisms (see the table) and placed in the medium.
Recovery obtained after 24 hours at 20-25 °C: **passed.**

TRANSYSTEM AMIES AND STUART CLEAR GEL AND WITH CHARCOAL	TRANSYSTEM AMIES AND STUART LIQUID	TRANSYSTEM CARY BLAIR
<i>N. gonorrhoeae</i> ATCC 43069	<i>N. gonorrhoeae</i> ATCC 43069	<i>C. jejuni</i> ATCC 33291
<i>B. fragilis</i> ATCC 25285	<i>S. pyogenes</i> ATCC 19615	<i>E. coli</i> ATCC 25922
<i>H. influenzae</i> ATCC 10211	<i>H. influenzae</i> ATCC 10211	<i>S. flexneri</i> ATCC 12022
/	/	<i>Y. enterocolitica</i> ATCC 9610

All reported results were obtained at time of Lot release.

This document has been produced electronically and is valid without a signature

Date of issue: **September 2021**