

Certificate of Conformity

VWR Catalogue Number	See List
Description	VWR®, Watch Glasses
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2026-07-09

Catalogue Number	Rim	Diameter [mm]	DIN	Linear expansion coefficient
216-2182	Fused	50	12 341	$3.3 \times 10^{-6} \text{K}^{-1}$
216-2183	Fused	60	12 341	$3.3 \times 10^{-6} \text{K}^{-1}$
216-2184	Fused	80	12 341	$3.3 \times 10^{-6} \text{K}^{-1}$
216-2185	Fused	100	12 341	$3.3 \times 10^{-6} \text{K}^{-1}$

Quality System Compliance

This certificate confirms that the product has been sourced from a supplier or manufacturer who has declared their Quality Management System (QMS) complies with **ISO 9001**.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Borosilicate glass 3.3.
Sterilization	Non-Sterile.
ATP Assay	Not Applicable.
DNase & RNase Free	Not Applicable.

BSE/TSE	Not Applicable.
Cytotoxicity	Not Applicable.
Non-Pyrogenic Statement	Not Applicable.
Latex Statement	Not Applicable.
BPA Statement	Not Applicable.
DEHP Statement	Not Applicable.
RoHS	Testing Not Applicable.
REACH Statement	Testing Not Applicable.
Storage Conditions	Ambient Temperature.
Shelf Life	Not Applicable.

Disclaimer: VWR's Quality Management System (QMS) is ISO 9001:2015 certified as a wholesale distributor of finished products, ensuring effective oversight of finished product handling, traceability, and distribution processes. VWR is not a manufacturer and does not engage in manufacturing, co-manufacturing or processing activities such as design, packaging, labeling, testing or quality release. Consequently, the supplier or manufacturer's QMS certification may differ from VWR's.

VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. This product is intended for research use only. It is not for use in diagnostic procedures, therapeutic applications, or any human or clinical use.

This document has been produced electronically and is valid without a signature.

Version History

Sl.no	Version number	Description	Month & Year
1	1.0	Created General CoQ using new layout	07-2026