



Air-Tite Products Co., Inc.
565 Central Drive
Virginia Beach, VA 23454 U.S.A.
www.air-tite.com 800-231-7762

B. Braun Melsungen AG

**Carl-Braun-Strasse 1
D-34212 Melsungen
<http://www.bbraun.com>
Tax No.: 2522502007**

Certificate of Compliance

Manufacturer's Product Code: NJ-4606728-02
Air-Tite's Product Code: AL10
Alternate Supplier's Product Code:

Product Description:
10 ML (12 ML) LL NORM-JECT BBRAUN AIR-TITE

Lot Number: 23C06C8
Original Document Number: 5313672384

Certificate Date: 2023-04-06
Manufacture Date: 2023-03-06
Expiration Date: 2028-03-01

Technical / Physical Test
According to the Quality Specification >>>>>>> Passes Test of:
DIN EN ISO 7886-1 (Single-use Syringes)
DIN EN ISO 8537 (Single-use Insulin Syringes only)

Chemical Test
According to the Quality Specification >>>>>>> Passes Test of:
DIN EN ISO 7886-1 (Single-use Syringes)
DIN EN ISO 8537 (Single-use Insulin Syringes only)
European Pharmacopeia Section 3.2.8 USP <6.6.1>

Assurance of Sterility >>>>>>> Sterile
Pyrogen Test >>>>>>> Pyrogen free in accordance with USP
Residual ETO >>>>>>> In accordance to ISO 10993-7

This certificate is a computer print-out and is valid without signature. Certificate has been recreated from information supplied by B. Braun Melsungen AG. The original manufacturer's certificate is on file.