


PDS No. 6291xx	<b>PRODUCT DATA SHEET</b>			Page 1 of 1
Revision 07	Contact Dish, 65 x 15 mm			 <b>greiner bio-one</b>
	Greiner Item-No. 6291xx			
Valid for Item-No.:	<b>629161</b> (sterile)	<b>629180</b> (sterile)		

1.	Description / Specification	
1.1	Description	Contact dish with square divisions, graduation, sterile. The special design allows direct contact of media and surface for hygiene monitoring. The germs that might be present stick to the media and colony units can be calculated due to the 10 x 10 graduation. <i>629161</i> : no vents <i>629180</i> : with vents
1.2	Dimensions	See customer drawings
1.3	Volume	-
1.4	Material / Resin	Dish: PS (Polystyrene), free from heavy metal Lid: PS (Polystyrene), free from heavy metal
1.5	Colour	Dish: clear Lid: clear
1.6	Sterilisation	SAL 10 <sup>-3</sup>
1.7	Quality Control	- <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="http://www.gbo.com/bioscience">www.gbo.com/bioscience</a> →Products →Literature →Technical Information→Chemical Resistance of Resins
2.6	Shelf life	5 years after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	20
3.2	Pieces / Box	600
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 06	Date 18 November 2013	Date 19 November 2013	Date 19 November 2013	
Date 03.03.2011	Name S. Kaelberer	Name Dr. T.Schreiber	Name A. Schulz	