

Certificate of Quality



Part Number	MFLX45504-00
Lot Number	NA
Part Description	LUER W/LOCKNUT M 1/16 PC 25/PK
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: Makrolon 2558550115
- The proper cleaning of this product for food applications is the responsibility of the end-user.
- CFR 177.1580 FDA Approved
- An independent biological laboratory performed United States Pharmacopeia Class VI Toxicity Tests, Pyrogens, Heavy Metals (Identified as Lead), Tissue Culture, and Hemolysis Tests on the proprietary compound and have passed according to protocol.
- Meets USP Class VI requirements.
- These products have been manufactured in an ISO-9001-registered facility.
- A lot control system is used to assure traceability to all raw materials.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.