

Certificate of compliance

Date of certificate: 2021 - 01 - 21

Page: 1/1

CORNING GOSSELIN SAS 123 Rue de Caëstre - CS40019 59529 HAZEBROUCK CEDEX **BORRE - FRANCE**

Product description

Square bottle 2.5L HDPE na graduated H262.7 Ø77 & screw cap with wad wh, STERILE SAL 10-3

Catalogue number *LC2500A-07*

Lot number

770993

Manufacturing date 2021/01/06

Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-3

Expiration date

2024/01/01

Expiration statement

This product is designed to have and expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

Cédric MAES

Plant Quality Manager, Borre

maesc@corning.com