

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE

Date of certificate : 2019 - 11 - 06

Page : 1/1

Product description

Container 90mL PP HingedCap Sterile 350/CS

Catalogue number

FL100-05

Lot number

718592

Lots : 718592

Manufacturing date

2019/09/05

Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10⁻³

Expiration date

2022/08/01

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.



Cédric MAES

Plant Quality Manager, Borre & Chorges

maesc@corning.com