

Certificate Of Analysis

Whatman™

Product Information

Product Number: 1822-110

Product Name: GR GF/C 11CM 100/PK

Lot Number/Serial Number: 17227340

Date of Manufacture: 4-Dec-2020

Quality Control

Test	Range	Results
Grammage,gsm	50-60	52
Thickness, um@53kPa	225-295	263
Porosity, S/300ml/1.0sq.in	13-27	26

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to ISO9001:2015 and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

Electronic signature

This document has electronically produced and is valid without a signature.

Version AD



www.gelifesciences.com
Global Life Sciences Solutions Operations UK Ltd.
Amersham Place Little Chalfont
Buckinghamshire HP7 9NA UK