



## VWR® TEST TUBE RACKS

As per the manufacturer, the below product meets the following criteria:

<b>VWR Catalog Number</b>	89215-722
<b>VWR Europe Catalog Number</b>	212-0385
<b>Lot Number</b>	V39-040120
<b>Description</b>	VWR Rack 13mm-36-PI Nylon Wht
<b>Date of Manufacture</b>	January 04, 2020
<b>Country of Origin</b>	India.

### Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2015 & ISO13485:2016 Quality Management System. Inspection records are reviewed and signed for product release.

### QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
Autoclaving	QWI-QC-07	General Insp. Level- I, AQL- 1.5	Pass
Dimensional	QWI-QC-23	General Insp. Level- I, AQL- 1.5	Within tolerance

### Product Specifications

- Material:** 30% Glass-filled Nylon with color masterbatch.
- BSE/TSE Statement:** The above-mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.
- Non-Pyrogenic:** Not Applicable.
- DNase & RNase Free:** Not Applicable.
- Cytotoxicity:** Not Applicable.
- Latex-Free Statement:** No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

Signed:



Jamie Ethier  
VP Global Quality  
VWR, Part of Avantor

**Date:** October 13, 2020