

---

## Certificate of Conformity

**Product:** Nunc Carrier plates, 24 Well Multidishes

**Date:** 2017-01 -20

**Page:** 1

**Item number:** 141004

**Lot number:** 149462

**Date of manufacture:** September 2016

**Date of expiry:** September 2021

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP <85> Bacterial endotoxins test	<0.5 endotoxins unit/ml	Complies
USP <88> Biological Reactivity test In Vivo, USP Class VI Test	USP current edition	Complies**
Monolayer formation - MRC-5 cell line	According to specification	Complies

\*\*Only performed once

Sterility is obtained through irradiation according to ISO 11137 Sterilization of health care products - Radiation, with SAL 10E-6.

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2008 Quality management system - Requirements and ISO 13485:2003 Medical devices - Quality management systems - System requirements for regulatory purposes.

Roskilde Site



Anne-Mette Reimann  
QA Manager