

Certificate of Quality



Part Number	MFLX04691-12
Lot Number	Q009791301
Part Description	PURISL PT TUBNG 3/16X11/32 25FT
Expiration Date	22NOV2024
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: Platinum-Cured Silicone
- The material used to manufacture this product is certified by testing conducted on representative samples and/ or attestations provided by approved vendors to meet the below criteria.
- USP Class VI - Yes
- Animal-Derived Component Free - Yes
- BSE/ TSE Compliant - Not applicable
- REACH - Yes
- RoHS - Yes
- ISO 10993 - Yes
- This product is not approved or intended for, and should not be used for medical, clinical, surgical or other patient-oriented applications.
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.