

# Certificate of Quality



Part Number	<b>MFLX06440-35</b>
Lot Number	<b>288921123</b>
Part Description	<b>TBNG MFLEX TYGON ELFL NO. 35 25FT</b>
Expiration Date	<b>18MAY2024</b>
Country of Origin	<b>USA</b>

## Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

## Product Specifications & Compliance

- Compound: Tygon E-LFL
- Product meets testing to Total Extractable in Rubber Articles Intended for Repeated Use Based on 21 CFR Part 177.2600.
- These products have been tested and have meet the acceptable limits as for requested tests of the European Pharmacopoeia (EP) Monograph: 3.2.9 - Rubber closures for aqueous parenteral preparations for powders and for freeze dried powders.
- A lot control system is used to assure traceability to all raw materials.
- This product has been qualified as meeting the criteria in the U.S.P for Class VI.
- Avantor is representing the Best if Used Date on the product label as Expiration Date on the certificate. Tubing specifications and/or characteristics may change past this date. It is solely up to the user to make this determination of suitability.

## Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

## Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.