



**Certificate of Conformance**

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

Product Part Number: MZSM56-14IR  
Lot Number: 008307-000  
Lot Quantity: 48 CASE(S)  
Date of manufacture: 06/17/2018  
Shelf Life: N/A  
Description: MZONE SNAPMOP NOVAPOLY, IRR'D (24/CS)  
Width: 4"  
Length: 14"  
Irradiation Reference Number: 0001426-00

06/21/2018

---

**Authorized Signature**  
**Quality Control Department**

**Date**

CC: With Product  
Customer File  
C of C File

# Certificate Of Processing

Prepared for MICRONOVA MFG INC

# STERIS

Gamma Process Run ID 109375A

<u>Product Code</u>	<u>Product Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
MZSM56-14IR	008307-000	48	CS

Other Information:

0001426-00

Processing Run Start Date/Time:	17-Jun-2018 01:49:02 pm	Approx. Downtime (hours):	0.00
Processing Run End Date/Time:	17-Jun-2018 09:43:50 pm		

Minimum Specified Dose (kGy):	40.0	Minimum Delivered Dose (kGy):	44.0
Maximum Specified Dose (kGy):	60.0	Maximum Delivered Dose (kGy):	57.3

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

### Signature Manifest

Reviewed and E-Signed By  
**Joanna Jimenez (QS & RC Technician)**  
Document Content Revision: 1

Signed On 6/19/2018 at 8:58 AM  
UTC / GMT Offset (hh:mm): -7:00

Processing Location:

STERIS  
1000 South Sarah Place  
Ontario, CA 91761  
Phone: 909-390-9942  
Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.