



Quality Certificate

VWR® SQUARE/OCTAGONAL MEDIA BOTTLES, POLYCARBONATE

As per the manufacturer, that these products meet the following Criteria.

Product Number:	89107-026
Lot Number:	090810
Description:	VWR BOTTLE PC 2L CLR SQ PK6
Production Date:	August 10, 2020
Country of Origin:	USA

Non-Toxin Leaching Resin: This product was manufactured in a class 10,000 cleanroom SABIC LEXAN 104 resin that is FDA compliant and complies with all requirements of the US Food, Drug and Cosmetic Act. It is deemed particle free.

Component Materials Toxicity: All component materials pass the requirements of the MEM elution cytotoxicity procedure utilizing WI 38 or MRC5 cell lines. The polycarbonate resin used is in compliance with Federal Food Additive Regulation 21 CFR 177.1580.

Heavy Metals: The resins used in the manufacture of Polycarbonate Bottle do not require reporting under the Superfund Amendment and Reauthorization Act (SARA) Title III Section 313. They would pass Toxicity Characteristic Leaching Procedure (TCLP) testing.

Lot Criteria: *The manufacturing lot was sampled, tested and released by Quality Assurance at the Manufacturer for the following characteristics:*

Signed:

Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: October 12, 2020