

## Certificate of Quality

**REF**

658175

**LOT**

E19033B5

Product description: CELL CULTURE FLASK, 250 ML, 75 CM<sup>2</sup>, PS,  
RED FILTER SCREW CAP, CLEAR,  
CELLSTAR® TC, STERILE, 5 PCS./BAG

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Free of detectable DNase/RNase/human DNA  
Representative product was tested by Real-Time-PCR.

Detection limits:

DNase 1.3 × 10<sup>-8</sup> Kunitz units  
RNase 5.0 × 10<sup>-11</sup> Kunitz units  
hDNA 5 pg

Free of detectable endotoxins (pyrogens)

Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>  
Detection limit: 0.03 EU/ml

Non-cytotoxic

Representative product was tested according to ISO 10993-5 and has been shown to be non-cytotoxic.

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

ISO 10993-5 "Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity"

ISO 11137 "Sterilization of health care products - Radiation"

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2023-03-04

**STERILE**

SAL 10<sup>-3</sup>

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2019-03-25

[www.gbo.com](http://www.gbo.com)