

# Certificate of Quality



Part Number	<b>MFLX96241-14</b>
Lot Number	<b>B238013147</b>
Part Description	<b>TUBING MFLEX STA-PURE NO. 14 12IN</b>
Expiration Date	<b>01JUN2026</b>
Country of Origin	<b>USA</b>

## Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

## Product Specifications & Compliance

- Compound: N/A
- Product has been sent to an independent laboratory(ies) (using GLP practices, 21CFR Part 58 where available) to undergo USP Class VI testing and cytotoxicity testing and were rated as "passing" in all tests.
- USP <88> Biological Reactivity Tests In Vivo, Class VI: PASS
- USP <87> Biological Reactivity Tests In Vitro: PASS
- This product is not intended for use in implantable patient applications (human or animal) including, but not limited to, medical and dental use.
- This product should never be sterilized by irradiation, including gamma and E-beam.
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

## Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

## Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.