

# Certificate of Quality



Part Number	<b>MFLX96407-25</b>
Lot Number	<b>29219658</b>
Part Description	<b>TUBING MFLEX SILICONE NO. 25 250FT</b>
Expiration Date	<b>01APR2027</b>
Country of Origin	<b>USA</b>

## Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

## Product Specifications & Compliance

- The proper cleaning of this product for food applications is the responsibility of the end-user.
- The product itself has not been tested for approval by the FDA or any other testing body.
- The polymer used in this product is compositionally compliant with 21 CFR,177.2600 - Rubber Articles intended for repeated use.
- Polymer sample has been tested to USP Class VI under GLP. The material did not produce biological response in rabbits.
- \*\*Users of this material are advised to review 21 CFR 177.2600 for details of extraction limitations and other requirements and assure themselves that the end product passes these requirements for FDA compliance under their specific processing conditions.
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

## Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

## Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.