

Certificate Of Analysis

Whatman™

Product Information

Product Number: 1827-090

Product Name: 934-AH 9.0CM 100/PK

Lot Number/Serial Number: 17227358

Date of Manufacture: 7-Dec-2020

Quality Control

| Test | Range | Results |
|----------------------------|---------|---------|
| Grammage,gsm | 57-71 | 64 |
| Thickness, um@3.5kPa | 390-480 | 408 |
| Porosity, S/300ml/1.0sq.in | 9-13 | 12 |

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to ISO9001:2015 and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

Electronic signature

This document has electronically produced and is valid without a signature.

Version AB



www.gelifesciences.com

Global Life Sciences Solutions Operations UK Ltd.

Amersham Place Little Chalfont

Buckinghamshire HP7 9NA UK