

# Certificate of Quality



Part Number	<b>MFLX06414-20</b>
Lot Number	<b>29827144</b>
Part Description	<b>TUBING VACUUM 1/4X5/8IN 10FT</b>
Expiration Date	<b>22JUN2025</b>
Country of Origin	<b>USA</b>

## Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

## Product Specifications & Compliance

- Compound: E-3603
- 21 CFR 176.170(c) under condition of use: E through H as described in Table 1 and 2. The food shall not be left in the tubing over 8 hours.
- Product has been manufactured in accordance with Good Manufacturing Practices.
- A lot control system is used to assure traceabilities to all raw materials.
- The proprietary compound has been tested: USP Class VI Toxicity Tests and have passed accordingly.
- Avantor is representing the Best if Used Date on the product as label as Expiration Date on the certificate. Tubing specifications and/or characteristics may change past this date. It is solely up to the user to determine suitability.

## Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

## Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.