

Certificate of Quality



Part Number	MFLX95687-16
Lot Number	29433174
Part Description	TUBING MFLEX PHARMED NO.16 500'
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound:PharmMed® BPT
- These products have been manufactured in an ISO-9001 registered facility.
- A lot control system is used to assure traceability to all raw materials.
- PharMed® BPT meets the requirements of Class VI USP <88>, Biological Reactivity Test, In Vivo - Clas VI andUSP <87> Biological Reactivity Tests, In Vitro # Cytotoxicity.
- REACH (EC 1907/2006) The product has been evaluated against Regulation (EC) No 1907/2006 of the European Parliament and of the Council of 18 December 2006 concerning the Registration, Article 33 (1). The product(s) referenced DO NOT CONTAIN any of the substances included in the REACH Authorization list (Annex XIV). In addition, the product(s) DO NOT CONTAIN any substances from the Substance of Very High Concern (SVHC) candidate list (Total 211 substances) in concentration above 0.1% weight by weight (w/w). This declaration is based on the raw material information provided by the manufacturer#s suppliers.
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.