

Eppendorf Certificate

Certificate of Purity

Product: Eppendorf Serological Pipet, 5 mL

Order-No.: 0030 127.714

Lot: L000242J

Products are manufactured in an class 8 clean room which meets the ISO 14644-1 specifications. Testing of the above described lot showed conformity within the limits of detection:

- | | | |
|-------------|--------------|--|
| • Human DNA | Method: | <i>real-time</i> PCR |
| | Limit: | < 2 pg; less than one human cell |
| • DNase | Method: | DNA Digestion |
| | Limit: | not detectable (LOD 1.0 x 10 ⁻⁶ Kunitz units) |
| • RNase | Method: | RNA Digestion |
| | Limit: | not detectable (LOD 1.0 x 10 ⁻⁶ Kunitz units) |
| • Endotoxin | Method: | Kinetic-turbidimetric LAL test (Ph. Eur.2.6.14) |
| | Limit: | < 0.01 EU/mL |
| • Sterility | Method: | Sterilization with SAL 10 ⁻⁶ by irradiation according DIN EN ISO 11137-1:2015-11 |
| | Test method: | USP, Ph. Eur. 8.0/ 2.6.12 |

Hamburg, April 2022



Joana Tziolis
Product Life Cycle Manager
Division Consumables



Monika Schneider
Vice President Global Quality Management &
Regulatory Affairs

Your local distributor: www.eppendorf.com/contact
Eppendorf SE · Barkhausenweg 1 · 22339 Hamburg · Germany
E-mail: eppendorf@eppendorf.com

ISO 9001
Certified

ISO 13485
Certified

ISO 14001
Certified