

## Certificate of Conformity

---

### Product Information

---

Product Number: 9913-2502

Product Name: Whatman™ Uniflo™ 25mm 0.22 PVDF S 45/PK

Lot Number: 200916-419

Sterility: GAMMA

Expiry Date: 16-Sep-2023

Date of manufacture: 16-Sep-2020

### Conformance & Quality systems statement

---

This is to certify that this product conforms to Cytiva specifications.

All products are manufactured via a Quality System certified to **BS EN ISO9001:2015** and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

This product has been gamma sterilised to achieve a SAL (sterility assurance level) of  $10^{-6}$  in accordance with the requirements of **ISO11137**

### Licenses and Trademarks

---

Whatman and Uniflo are trademarks of Global Life Sciences Solutions USA LLC or an affiliate doing business as Cytiva.

Cytiva and the Drop logo are trademarks of Global Life Sciences IP Holdco LLC or an affiliate.

© 2020 Cytiva.

### Electronic signature

---

This document has been electronically produced and is valid without a signature.

Version AB

cytiva.com