

Certificate of Quality



Part Number	MFLX06407-72
Lot Number	30105023
Part Description	TUBING TYGON 1/16X3/16IN 50FT
Expiration Date	07AUG2025
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: E-3603
- Product has been manufactured in accordance with Good Manufacturing Practices.
- A lot control system is used to assure tracing to all raw materials.
- The proprietary compound has been tested: USP Class VI Toxicity Tests and have passed accordingly.
- Avantor is representing the Best if Used Date on the product as label as Expiration Date on the certificate. Tubing specs. and/or characteristics may change past this date. It is solely up to the user to make this determination of suitability.
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.