

Certificate of Quality



Part Number	MFLX30800-16
Lot Number	314203
Part Description	LUER ADAPTER M 1/16 PP 25PK
Expiration Date	01JAN2032
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: Polypropylene
- Product has been visually inspected by the manufacturer to a 0.65 AQL. Sample size is based on ANSI/ASQZ 1.4 Table II-A Single sample plan for normal inspection.
- Meets regulatory requirements RoHS / RoHS 2 / RoHS 3. The material supplier has declared that this material complies with 2011/65/EC and (EU) 2015/863.
- Manufacturer has not directly tested the product for restricted or banned substances in the above directives.
- Resin manufacturer has certified the plastic resin to USP Class VI.
- TSE/BSE No tallow or any other animal derived materials used in manufacturing or formulation of this product.
- Product has not been tested to ANY specific regulatory requirement, therefore the enduser must make their own determinations ensuring the use of the products are safe, lawful, and suitable for their intended applications.
- Proper cleaning of this product prior to first use is recommended.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.