

Certificate of Quality



| | |
|-------------------|--|
| Part Number | MFLX96400-25 |
| Lot Number | 28593204 |
| Part Description | TUBING MFLEX SILICONE NO. 25 25FT |
| Expiration Date | 31DEC2049 |
| Country of Origin | USA |

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: SM2150A
- All ingredients used in these products are approved under the Code of Federal Regulations 21, subchapter B, Part 177.2600 - Rubber Articles intended for repeated use in contact with food (Food-Grade).
- These products have been manufactured in accordance with Good Manufacturing Practices.
- A lot control system is used to assure traceability to all raw materials.
- An independent biological laboratory performed United States Pharmacopeia Class VI Toxicity Tests, Pyrogens, Heavy Metals (Identified as Lead), Tissue Culture, and Hemolysis Tests on the proprietary compound and have passed according to protocol.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.