

Certificate of Quality

VWR Catalogue Number	See List
Description	VWR® Verification Thermometers
Country of Origin	Manufactured in USA
Date of Issue (yyyy-mm-dd)	2022-04-12

Thermometers

VWR Catalogue Number	Dimensions	Range	Length [mm]	Accuracy	Divisions
89095-678	Overall: 1.5" x 1.5" x 4.9" (125mm) H Bottle 1.5" x 1.5" x 2.44"H, 30ml	-5 to 20°C	125	± 0.75°C	0.5°C
75831-368	Overall: 2" x 2" x 4.9" (125mm) H	-5 to 20°C	125	± 0.75°C	0.5°C
75831-366	Overall: 2" x 2" x 7.1" (180mm) H	-2 to 10°C	180	± 0.4°C	0.1°C
89095-670	Length: 5.3" Bottle: 30ml	-30 to 1°C	135	± 0.75°C	0.5°C
89095-684	Bottle 1.5" x 1.5" x 2.44"H, 30ml Overall: 1.5" x 1.5" x 5.7" H	+10 to 45°C	125	± 0.5°C	0.5°C
89095-686	Overall: 1.5" x 1.5" x 5.3" (135mm) H Bottle 1.5" x 1.5" x 2.44" H	+15 to 50°C	135	± 0.75°C	0.5°C
89095-690	Overall: 1.5" x 1.5" x 5.7" (144mm) H Bottle 1.5" x 1.5" x 2.44"H	+20 to 130°C	145	± 1.5°C; ± 2°C above 100°C	1°C
89095-692	Overall: 1.5" x 1.5" x 5.7" 145mm) H Bottle 1.5" x 1.5" x 2.44"H	+50 to 200°C	145	± 2.5°C; ± 5°C above 100°C	2°C
89095-676	Overall: 1.5" x 1.5" x 4.9" H Bottle 1.5" x 1.5" x 2.44"H, 30ml	-5 to 15°C	125	± 0.75°C	0.5°C

89095-680	Bottle 1.5" x 1.5" x 2.44"H, 30ml Overall: 1.5" x 1.5" x 5.7" (135mm) H	-90 to 25°C	145	± 1.5°C; ± 2.5°C below - 50°C	1°C
-----------	--	----------------	-----	-------------------------------------	-----

VWR CATALYST Re-Calibration Service (Service Only)

VWR Catalogue Number	Description
10186-012	Customer Specified 1 Point Re-Calibration (0°C - 100°C)
10186-014	Customer Specified 2-3 Points Re-Calibration (0°C - 100°C)

Quality System Compliance

Products are manufactured under the [ISO 9001:2015 & ISO/IEC 17025:2017](#) standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable product specifications.

Product Specifications

Material	Glass Beads, individually serialized, Lead-Free Glass Thermometer with Blue Organic Liquid Fill; Traceable to NIST, Traceable to DAkkS
Sterilization	Not applicable.
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.
BSE/TSE	Not applicable.
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	Not applicable.
BPA Statement	Not applicable.
DEHP Statement	Not Applicable.
RoHS	Not Applicable.
REACH Statement	Not Applicable.

Storage Conditions Store at ambient warehouse temperatures, low humidity, protected from direct sunlight.

Shelf Life Product has no expiry date.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.