

Certificate of Quality



Part Number	MFLX95718-18
Lot Number	28233955
Part Description	PUMP TUBING C-FLEX 0.51 MM ID 50 FT
Expiration Date	27OCT2026
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: R70-001 Medical Grade C-Flex
- An independent biological laboratory performed United States Pharmacopeia Class VI Toxicity Tests, Pyrogens, Heavy Metals (Identified as Lead), Tissue Culture, and Hemolysis Tests on the proprietary compound and have passed according to protocol.
- Meets USP Class VI requirements.
- This product has been manufactured in accordance with Good Manufacturing Practices.
- A lot control system is used to assure traceability to all raw materials.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.