

## Certificate of Quality

**REF**

664161

**LOT**

E20113BD

Product description: PETRI DISH, 100/20 MM, PS, CLEAR, WITH VENTS,  
STERILE, 15 PCS./BAG

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Non-cytotoxic

Representative product was tested according to ISO 10993-5 and has been shown to be non-cytotoxic.

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

ISO 10993-5 "Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity"

ISO 11137 "Sterilization of health care products - Radiation"

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2025-10-26

**STERILE**SAL 10<sup>-3</sup>

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2020-11-20

[www.gbo.com](http://www.gbo.com)