PDS No. 6561xx	**PRODUCT DATA SHEET**	Page 1 of 1
Revision 04	Lid for Microplates, PS Greiner Item-No. 6561xx	greiner bio-one

Valid for Item-No.:	**656101**	**656161** (sterile)	**656170**	**656171** (sterile)	**656190**
	656191 (sterile)				

1.	**Description / Specification**	
1.1	Description	PS lid for Microplate, single position *656101*: high profile *656161*: high profile, sterile *656170*: high profile, condensation rings for 96 well plate *656171*: high profile, condensation rings for 96 well plate, sterile *656190*: low profile *656191*: low profile, sterile
1.2	Dimensions	*656101, -161, -170, -171*: length: 127,5 mm (+/- 0,2 mm) width: 85 mm (+/- 0,2 mm) curvature: +/- 300 µm *656190, -191*: length: 127,35 mm (+/- 0,15 mm) width: 84,8 (+/- 0,15 mm) curvature: +/- 300 µm
1.3	Volume	N/A
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	*656161, -171, -191*: SAL 10^{-3} *656101, -170, -190*: no
1.7	Quality Control	- Raw Material-Control: physical testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	- For single use only - *656190, -191*: the covered microplates can be easily gripped on the side in automated processes

2.	**Features**	
2.1	Basic features	All Item-No.: free of detectable DNase/RNase, human DNA and pyrogens *656161, -171, -191*: contents non-cytotoxic
2.2	Temperature range	-20℃ to +60℃
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: www.gbo.com/bioscience →Products →Literature →Technical Information →Chemical Resistance of Resins
2.6	Shelf life	*656101, -170, -190*: n/a *656161, -171, -191*: 5 years after month of production
2.7	Other Information	-

3.	**Packaging**	**656101, -170**	**656161, -171**	**656190**	**656191**
3.1	Pieces / Bag	1	1	20	20
3.2	Pieces / Box	400	400	200	200
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)			
3.4	Other Information	-	Certificate of Quality	-	Certificate of Quality

4.	**Other Information**

Data Sheet subject to change without notice!

Prior Issue	**Drawn**	**Approved**	**Released**
Revision 03	Date 20 September 2011	Date 21 September 2011	Date 21 September 2011
Date 14.12.2009	Name S. Kaelberer	Name Dr. R. Heller	Name A. Schulz

CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietory to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.