



## Quality Certificate

---

### VWR® SCINTILLATION VIALS, BOROSILICATE GLASS, WITH SCREW CAPS

---

As per the manufacturer, the below product meets the following criteria:

<b>North American Catalog No:</b>	66022-065
<b>Lot Number:</b>	5-3A-3
<b>Description:</b>	VWR Vial Scint Gls 20ml Cs500

The product listed above was manufactured, using glass that conforms to the ASTM E438 Specifications for glass in Laboratory Apparatus for:

***Type I, Class "A" Borosilicate 33 Expansion Glass – Clear***

Which also conforms to the most current revision of the United States Pharmacopoeia USP for Type I Borosilicate glass and it also meets the USP Limit for Glass Grain and Arsenic Extraction. Glass material will meet the requirements of the most current revisions of the European Pharmacopoeia and Japanese Pharmacopoeia for Type I, or Hydrolytic Glass.

The product was inspected and accepted in accordance with the manufacturer's ISO-9001:2015 Quality Management System and corresponding inspection and monitoring procedures to an Acceptance Quality Limit of 1.0 per ANSI/ASQC Z1.4.

**Signed:**

Jamie Ethier  
VP Global Quality  
VWR, Part of Avantor

**Date:** January 04, 2020