

## Certificate of Quality

**REF**

756071

**LOT**

E20113UB

Product description: STRIP PLATE, 6x16, PS, F-BOTTOM,  
CLEAR, MICROLON®, HIGH BINDING, 10 PCS./BAG

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Free of detectable DNase/RNase/human DNA  
Representative product was tested by Real-Time-PCR.

Detection limits:

DNase  $1.3 \times 10^{-8}$  Kunitz units  
RNase  $5.0 \times 10^{-11}$  Kunitz units  
hDNA 5 pg

Free of detectable endotoxins (pyrogens)

Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>  
Detection limit: 0.03 EU/ml

Binding Property

MICROLON® coefficient of variation < 5% (IgG ELISA / transmission measurement)  
FLUOTRAC# coefficient of variation < 10% (IgG ELISA / fluorescence measurement)  
LUMITRAC# coefficient of variation < 10% (IgG ELISA / luminescence measurement)

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2024-11-12

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2020-12-09

[www.gbo.com](http://www.gbo.com)