

TECHNICAL DATA SHEET
Latex, Non-Sterile Medical Examination Glove: ISO 11193-1 - Type 1
Intended Use

Non-sterile, single use, medical examination gloves are intended to prevent transmission of a wide variety of diseases to both patients and health care personnel. Specifically, gloves are intended to act as a barrier between the healthcare provider and patients to prevent contamination between patient and caregiver.

ISO 11193-1 - Type 1			
Characteristic			Acceptable Limit
Dimension & Tolerance	Width (mm)	Extra-Small (XS)	≤ 80
		Small (S)	70 - 90
		Medium (M)	85 - 105
		Large (L)	100 - 120
		X-Large (XL)	≥ 110
	Length (mm)	XS & S: ≥ 220 Others: ≥ 230	
Thickness (mm)	Palm	≥ 0,08	
	Finger	≥ 0,08	
Physical Requirement	Before ageing	Force at Break, N	≥ 7,0
		Elongation at Break, %	≥ 650%
	After ageing	Force at Break, N	≥ 6,0
		Elongation at Break, %	≥ 500%
Powder Residue	ASTM D 6124	≤ 2,0 mg/glove	
Latex Protein	ASTM D5712	≤ 50 µg/dm ²	
Description			Ambidextrous.

Additional Information

Storage Condition	Country of Origin
Store in a cool, dry place. Avoid excessive heat (40°C, 104°F). Shield open box from direct sun and fluorescent light.	Malaysia

Appendix 1 List of Item Codes and Attributes

Appendix 1 (List of items)

AMD Medicom Inc.

Item Code	Description	Packaging Information	Size	Color	Specific Attribute	Treatment	Marketed in USA as	Marketed in Canada as
9992	Medicom AssureTouch Powder-Free Latex Medical Examination Gloves	100 Gloves/Box, 10 Boxes/Case, 1000 Gloves/Case	A = X-small, B = Small, C = Medium, D = Large, E = X-Large	Natural	Textured	Non-Sterile	Not Marketed In USA	Medical Device Class 2