

## CERTIFICATE OF CONFORMITY & CERTIFICATE OF ANALYSIS

Brand	TouchNTuff®
Size(s)	M (7.5 - 8)
Material	Nitrile
Style	93-700
Short Description	Sterile Nitrile Disposable Glove
Lot Number	2211SW
Batch Number	22117008SW
Manufacturing Date	2022-11
Expiration Date	2025-10

### DECLARATION OF CONFORMITY

Ansell certifies that the above product is manufactured in accordance with Ansell product specifications and is in conformity with the following standards or normative documents:

Standard/Normative Document	Result
FDA 21 CFR 177-2600	Conforms
Free from BSE/TSE (EMEA/410/01)	Conforms
ISO 11137-1	Conforms
PPER (EU) 2016/425	Conforms
Reach Regulation	Conforms

### DECLARATION OF ANALYSIS

Property	Test Method	Specification	Result
Freedom from Holes - Final Release	EN455-1/ASTM D5151	G1, AQL 1.5	Conforms
Dimension Test	ASTM D3767	S-2, AQL 4.0	Conforms
Glove Length (mm)	ASTM D3767	Min. 300	Conforms
Palm Thickness (mm)	ASTM D3767	0.11-0.15	Conforms
Finger Thickness (mm)	ASTM D3767	0.14–0.18	Conforms
Physical Properties	ASTM D412	S-2, AQL 4.0	Conforms
Tensile Strength (MPa)	ASTM D412	min. 14	Conforms
Ultimate Elongation (%)	ASTM D412	min. 500	Conforms
Liquid Particle Count 0.5-20 µ (particles/cm <sup>2</sup> )	IEST-RP-CC005.4	< 1200	Conforms
Protein Content (µg/g)	EN455-3	Not applicable	Not applicable

Electronically Signed by:

**Amy Jones**

Manager, QA

Date: 29 March 2023



For additional information visit us at [www.ansell.com](http://www.ansell.com), or call us at:

**Europe, Middle East & Africa Region**

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

**Australia**

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

**Latin America & Caribbean Region**

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

**North America Region**

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

**Asia Pacific Region**

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699