



Becton Dickinson Medical (S) Pte Ltd
30 Tuas Avenue 2
Singapore 639461 SG

CERTIFICATE OF COMPLIANCE

Page: 1 of 1

Product Name : NEEDLE ECLIPSE 22X1-1/2 RB
Catalog Number : 305763 **Manufacture Date** : 2015/05/02
Batch Number : 5106267
Expiration Date : 2020/04/30

This is to certify that the product listed above was manufactured in accordance with Becton Dickinson Medical (S) Pte Ltd Product Specification no. SP100123 current revision.

Quality Control steps were carried out accordingly to established quality system at the manufacturing site.

The above products conform to the following:

ISO 7864 Standard

Sterile Hypodermic Needles for single use.

Sterility

EN 556-1 Sterilization of medical devices- Requirements for Medical devices to be designated "STERILE".

Requirements for terminally sterilized medical devices. The above referenced lot is certified sterile and the process of sterilization has been validated following ISO 11135.

All product which are labeled as sterile and released for sale by Becton Dickinson Medical (S) Pte Ltd are certified to be sterile as long as the package is unopened and undamaged.

Pyrogenicity

All products which are labeled as non-pyrogenic and released for sale by Becton Dickinson Medical (S) Pte Ltd have been tested and found to be non-pyrogenic.

GMP/Manufacturing

All Products of Becton Dickinson Medical (S) Pte Ltd are manufactured in accordance with the medical device's GMP requirements. The devices are manufactured in facilities registered* with US FDA; the devices are also listed* with US FDA. *per 21 CFR 807

Lim Siong Pew
QA Manager

Signature Date: 2015/06/19